



NC DEPARTMENT OF  
**HEALTH AND  
HUMAN SERVICES**  
Division of Public Health

ROY COOPER • Governor

MANDY COHEN, MD, MPH • Secretary

MARK T. BENTON • Assistant Secretary for Public Health

Division of Public Health

SCOTT M. SHONE, PhD, HCLD (ABB) • Laboratory Director, State  
Laboratory of Public Health

## Memo

**To:** Evelyn Foust  
Communicable Disease Branch Head

**From:** Angela Bradley, MLT (ASCP)  
Medical Laboratory Technologist

**CC:** Pete Moore  
Ron Higginbotham  
NCSLPH Lab Improvement Staff  
Local Health Department Lab Managers

**Date:** January 17th, 2020

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

**Modified Thayer Martin II:** Lot Number: 9347694 Expiration Date: **4/9/2020**  
Along with **Bio-Bag™ Type C:** Lot Number: 8129674 Expiration Date: **01/31/2020**  
and with **Bio-Bag™ Type C:** Lot Number: B02B133M Expiration Date: **10/31/2020**

### Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth	<i>S. epidermidis</i> ATCC# 12228 – No Growth
<i>N. gonorrhoeae</i> ATCC# 43070 - Growth	<i>P. mirabilis</i> ATCC# 43071 – * Growth
Sterility Check - Acceptable	Visual Check - Acceptable

\*While performing quality control testing on the Modified Thayer Martin II agar plates, there was growth of 1 colony of *Proteus mirabilis* (Oxidase-negative, Gram negative rod) at 24 hours. Please pay attention to possible growth of other organisms, especially at the 24-hour mark. Follow your laboratory's procedure for the presumptive identification of *Neisseria gonorrhoeae*.

**Lot number failed:** None

BD and Fisher Scientific contacted: January 17<sup>th</sup>, 2020

\*\*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at [angela.bradley@dhhs.nc.gov](mailto:angela.bradley@dhhs.nc.gov). Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

[www.ncdhhs.gov](http://www.ncdhhs.gov) • <http://publichealth.nc.gov> • [slph.ncpublichealth.com](http://slph.ncpublichealth.com) • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER